

Dressing Application Guide

Invia® Foam Dressing Kit with FitPad



1. Thoroughly clean and debride the wound.

Apply skin prep to protect the peri-wound skin.



2. If required, apply a non-adherent contact layer prior to placing the foam into the wound.



Cut the foam to fit the size and shape of the wound. Do not cut directly over the wound.



4. Rub the edges of the foam to remove any loose particles.



5. Place the foam into the wound cavity. Do not tightly pack or force foam into areas of the wound. Do not allow the foam to overlap onto intact skin.



6. Apply transparent film. Partially peel back one side of layer 1 and place the adhesive side down. Remove the remaining side of layer 1.



7. Now remove the transparent backing layer labeled 2.



8. Remove the perforated silver colored tab on the side. The film should extend 3–5 cm beyond the wound margin to facilitate adequate seal.



9. Select
appropriate
location for the
suction pad
(FitPad) to be
applied. Pinch
the film and cut
a small hole
(approx. 1 cm).



 Peel the backing off the FitPad to expose the adhesive.



11. Place the
FitPad
centered over
the previously
cut hole in the
dressing.
Press firmly
for adherence.



12. Attach the
dressing tubing
to the pump
tubing by
pushing the
Quick-connector together until
you hear a click.



The Invia Foam Dressing Kit with FitPad is intended to be used in conjunction with the Invia Motion and Invia Liberty Negative Pressure Wound Therapy Systems (NPWT).



Invia NPWT instructions advise 24 hour therapy without interruption. If therapy is discontinued for more than 2 hours, the dressing should be replaced and therapy restarted by a healthcare professional.





Dressing Application Guide

Invia® Gauze Dressing Kit with FitPad



Thoroughly clean and debride the wound.
 Apply skin prep to protect the peri-wound skin.



2. If required, apply a non-adherent contact layer prior to placing the gauze into the wound.



Cut the gauze

 to fit the size
 and shape of the
 wound. Do not
 cut the gauze
 directly over the
 wound.



 Saturate the gauze with the saline provided.



5. The gauze should fit loosely into the wound bed. Do not tightly pack or force gauze into areas of the wound.



6. Apply
transparent
film. Partially
peel back one
side of layer 1
place the
adhesive side
down. Remove
the remaining
side of layer 1.



Now remove the transparent backing layer labeled 2.



8. Remove the perforated silver colored tab on the side. The film should extend 3–5 cm beyond the wound margin to facilitate an adequate seal.



9. Select
appropriate
location for the
suction pad
(FitPad) to be
applied. Pinch
the film and cut
a small hole
(approx. 1 cm).



 Peel the backing off the FitPad to expose the adhesive.



11. Place the FitPad centered over the previously cut hole in the dressing. Press firmly for adherence.



12. Attach the dressing tubing to the pump tubing by pushing the Quick-connector together until you hear a click.



The Invia Gauze Dressing Kit with FitPad is intended to be used in conjunction with the Invia Motion and Invia Liberty Negative Pressure Wound Therapy Systems (NPWT).



Invia NPWT instructions advise 24 hour therapy without interruption. If therapy is discontinued for more than 2 hours, the dressing should be replaced and therapy restarted by a healthcare professional.



This Quick Card does not replace the instructions for use REF 200.9065.

Please contact us or your local Medela representative for details.



Medela AG Lättichstrasse 4b 6340 Baar, Switzerland www.medela.com USA

Medela LLC 1101 Corporate Drive McHenry IL 60050, USA Phone +1 877 735 1626

Fax +1 815 307 8942 info-healthcare@medela.com

Canada

Medela Canada Inc. 4160 Sladeview Cres., Unit #8 Mississauga, ON L5L 0A1, Canada Phone +1 800 435 8316 Fax +1 800 995 7867 info@medela.ca Local contact:

© Medela AG/201.0869/2021-06/B Medela US/1908565/B